

Important information about your 1ml ENFit syringes


Please keep this letter with your syringe supplies.

Dear Patient, Parent or Carer,

We are writing to let you know about a printing error affecting certain batches of the Enteral-ISO SAF Home 1ml Low Dose ENFit Single Patient Use Syringe supplied by GBUK.

What might you notice?

You may see a single-use symbol (like on the picture below) printed on the barrel of some syringes in your delivery. This symbol has been printed in error and should not be there.



What does this mean for you?

- The syringes remain safe to use as usual.
- The printing error does not affect how the syringe works, its safety or the materials it is made from.
- The measurement markings on the syringe remain correct.
- You do not need to return, replace or throw away the affected syringes because of this printing error.

Please continue to use and care for the syringes in the usual way, following your existing guidance.

Do you need to do anything?

No. GBUK has confirmed that the syringes remain safe and suitable for normal use, so no action is needed.

If you have a question

Please contact Nutricia Homeward using your usual contact details. Our team will be happy to support you.

Kind regards,

Nutricia Homeward

You're receiving this information because your healthcare professional registered you for the Homeward service, and we have your details on file. This communication contains important information related to your Homeward service.

This communication is sent by Nutricia Limited (company number 00275552), registered at Newmarket House, Newmarket Avenue, White Horse Business Park, Trowbridge, Wiltshire, BA14 0XQ.